



UK Standards for Microbiology Investigations

Review of users' comments received by Working group for microbiology standards in clinical virology/serology

Q 4 Good practice when performing molecular amplification assays





"NICE has renewed accreditation of the process used by **Public Health England (PHE)** to produce **UK Standards for Microbiology Investigations**. The renewed accreditation is valid until **30 June 2021** and applies to guidance produced using the processes described in **UK standards for microbiology investigations (UKSMIs) Development process, S9365', 2016**. The original accreditation term began in **July 2011**."

Issued by the Standards Unit, Microbiology Services, PHE RUC | Q 4 | Issue no: 2 | Issue date: 19.02.18 Page: 1 of 17

First consultation: 11/08/2017 – 25/08/2017

Version of document consulted on: Q 4dq+

Proposal for changes

Comment number	1			
Date received	14/08/2017	Lab name	PHE Porton	
Section	Multiple			
Comment				
laboratory practice as t	his has legal me	ut) of this document to remove eaning in a quality context. Per m GLP will attract an Audit to t	haps rename it as	
Evidence				
MHRA website http://w	ww.legislation.g	ov.uk/uksi/1999/3106/regulatio	on/3/made	
Financial barriers				
Not completed.				
Health benefits				
Failure to address this the reputation of PHE.	issue may resul	t in an audit failure by MHRA v	vhich would affect	
Recommended	NONE			
action This issue has already been addressed in the document prior to consultation. The version that went to consultation had the updated title "Good practice when performing molecular amplification assays" and amendments made throughout the document. The amendment table on page 4 has the updated information.				

Comment number	2		
Date received	15/08/2017	Professional body	ACM ,UKCVN & UKAS assessor
Section	6 Quality issues		
Comment			

Although I accept that the document states 'it is important to demonstrate that assays are performing consistently....'I am surprised that there is no mention of demonstration of spectral calibration in real time PCR or evidence that amplification platforms are meeting the required temperature calibration as required by ISO 15189 -both being critical for the

procedures to obtain reliable results. The use of external biological controls can be used as a surrogate as long as the data is collected and analysed appropriately.

Evidence

Not completed.

Financial barriers

Some small expense.

Health benefits

No.

Recommended	ACCEPT
action	This comment will be considered in the UK SMI Q1 document: Evaluations, Validations and Verifications of Diagnostic Tests when it comes up for review. A change request has been raised to that effect.

Com	nent number	3		
Date received		22/08/2017	Lab name	National Infection Service, Public Health England
Section	Section Multiple			
Com	nent			
a. Amend throughout all the hpa.org.uk website addresses to reflect the 01/04/2014 change from HPA to PHE. The links are correctly redirected as a legacy function, but need updating to their new final destination URLs in the reviewed document.			s a legacy function,	
Sever	al paragraphs to b	be added to the	document:	
b. Add para at the end of Introduction: Next-Generation Sequencing (NGS) is emerging as a powerful new diagnostic technique. Almost all applications involve one or more PCR amplification steps, so the mitigating practices described hereir also apply to NGS applications. Additional contamination risks are presented by NGS, particularly surrounding the use of adaptor/index molecules throughout an NGS workflow. Further steps to control these risks will be essential, but are beyond the scope of this document.				
C.	c. Organisation of Work, para 6: When discussing UV decontamination, it is imperative to state that weekly monitoring of bulb strength is necessary to ensure sufficient decontamination effect			
d.	d. Add para at the end of 1.1 Organisation of Work: It is very important that the are where specimens are received into the testing facility remains PCR 'clean', with no cloned or PCR-amplified material being handled. If such material is received by the testing laboratory, a separate, dedicated area for processing should be available, with its own equipment, lab coats, etc.		PCR 'clean', with aterial is received	
e. Title of 2.1: change to Physical and Temporal Separation of Pre-PCR and Post PCR Assay Stages		re-PCR and Post-		

- f. Add para at the end of 2.1: Where multiple overlapping PCR-based assays are being performed, in order to minimise the possibility of adventitious transfer of downstream material into clean areas, it is advantageous to perform clean tasks early in the working day, and 'dirty' tasks later, once the clean tasks have been completed, in order to minimise the possibility of adventitious transfer of downstream material into clean areas.
- g. 2.2, para 2: Change practise (verb) to practice (noun)
- h. Amend bullet point 1, section 4 Selection of Controls: A positive amplification control: this should normally be an extract that amplifies weakly but consistently within an acceptable range. A decline in assay performance may not be detected when using a high copy-number positive as this may still give a signal. Use of a strong positive is also an unnecessary risk as it can be a possible reservoir of contamination.

Evidence

Not completed.

Financial barriers

No.

Health benefits

The entire scope of this and other SMI documents is to ensure accurate conduct of diagnostic pathology testing, in this case using nucleic acid amplification techniques. Consequently every component of the SMI has health benefits, side effects and risks that might affect the development of this UK SMI. I'm not sure this question is appropriate!

Recommended	a.	NONE
action		This has already been addressed in the updated version of the document sent out on consultation. There is a possibility that the version of the document that was looked at is the version under review on the gov.uk website.
	b.	ACCEPT
		Information on NGS has been added to the document accordingly.
	C.	ACCEPT This has been updated in the document accordingly.
	d.	ACCEPT
		This has been updated in the document accordingly.
	e.	NONE
		It was agreed by the Working Group members that the title heading be kept as the same.
	f.	NONE

It was agreed by the Working Group members that the comment is not practical in laboratories and so this should not be mentioned in the document.
g. NONE
This has already been updated in the document accordingly.
h. ACCEPT
This has already been updated in the document accordingly.

Comment number	4			
Date received	22/08/2017	Lab name	PHE Public Health Laboratory Birmingham	
Section	1.1; 3; 4			
Comment				
areas should not that have been ir	be taken into c contaminated	ence Workbooks that have lean PCR areas. To workbo areas should not be taken i kely item in a diagnostic lab	oks or worksheets nto clean PCR areas.	
b. 3: I suggest changing the sentence Mastermixes should be subjected to minimal thawing and put on ice as soon as possible to Mastermixes should be subjected to minimal thawing and put on ice or a cooling block as soon as possible. Not many busy diagnostic labs will still be using ice as cool blocks are so much cleaner and more convenient.				
c. 4: I suggest changing the sentence Demonstration of the internal control sequence by PCR in a duplexed reaction with the target to Demonstration of the internal control sequence by PCR in a multiplexed reaction with the target In many labs, single target PCRs are not very common and the internal control will be part of a triplex or quadruplex assay.				
Evidence				
Experience.				
Financial barriers				
No.				
Health benefits				
No.				
Recommended action a. NONE This has already been updated using the word "documentation".			ng the word	

b. NONE

This has already been updated in the document.
c. ACCEPT
This has been updated in the document accordingly.

Comr	nent number	5			
Date received		24/08/2017	Lab name	Microbiology Dept, Belfast Trust	
Section	on	Multiple section	ns		
Comr	nent				
amplif layout This d refere	ication assays SM overall is not fluid locument details C nce to newer mole	II. While this is there are mul LP for molecul cular working s	n the GLP when performing m a very informative and useful of tiple sections where the order ar assays with a traditional set set ups where total enclosed a specimen processing, extract	locument I think the seems mixed up. up- there is no utomated platforms	
Evide	nce				
a.	Section 1.1: This	is a bit confuse	ed- needs reorganised.		
i	i. Organisation of work: 'Avoid entering pre-amplification rooms immediately after working in rooms where products, cloned material etc are handled'- This is unrealistic- how does one define immediately?				
ii. Also there are specific precautions in place as detailed in section 2.6 to prevent/mitigate the risks of contamination.		ection 2.6 to			
iii. 'All new members must be trained in use of PCR facilities'- this could more specific		es'- this could be			
iv	 'For reverse t specific preca 		ecific precautions are necessar	y' What are these	
b. Section 2.1: It is essential to have some reference in section 2.1 to automated start to finish systems - these are becoming much more commonplace in molecular UK laboratories and the SMI needs to address the GLP when using these systems.		onplace in			
 Section 2.2: 'The unidirectional Workflow'- may be better to include this in '1. Organisation of Work' section. 		lude this in '1.1			
d.	Section 2.3: 'the room	PCR machine r	room' should be changed to PC	CR amplification	
'Extraction and P for these- labs us		CR setup room	ktraction room' This should b many UK labs no longer ha sh automated high throughput setup) must be in the same ro	ive separate rooms systems (sample	

edited' this	f. Section 2.5: 'Individual users' PCR programs in the thermalcyclers should not be edited' this should be deleted as this is a policy that should be decided at local individual level.				
This should be c	Section 3: 'Mastermixes should be subjected to minimal thawing and put on ice' This should be changed to ' Mastermixes should be subjected to minimal thawing and handled as per manufacturer instructions'				
	tion of controls A positive amplification controlwithin an e' should be changed to 'within a locally defined range'				
i. Section 4: Selec	tion of controls:				
	Extraction controls' This may need rephrased- seems a bit - are we really talking about process controls??				
	mportant to include that a control of the whole process using a PCR reaction is acceptable.				
j. Section 5: 'Regular environmental swabbing is recommended' should be changed to 'Environmental swabbing can be useful' Environmental sampling can certainly be useful but recommending it as a regular necessity in a molecular laboratory can result in difficulties for individual laboratories- Do you test for every PCR target? How and where to swab? How often? What to do if positive? Is there a document that refers to best practice for this?					
k. References:					
Reference 7: Is i	it appropriate to reference a company in this SMI?				
Financial barriers					
No.					
Health benefits					
No.					
	a. i. ACCEPT				
No.	a. i. ACCEPT This has been addressed in the document accordingly.				
No. Recommended					
No. Recommended	This has been addressed in the document accordingly.				
No. Recommended	This has been addressed in the document accordingly. ii. NONE This has already been addressed in the document				
No. Recommended	This has been addressed in the document accordingly. ii. NONE This has already been addressed in the document accordingly.				
No. Recommended	This has been addressed in the document accordingly. ii. NONE This has already been addressed in the document accordingly. iii. NONE This has already been addressed in the document				
No. Recommended	This has been addressed in the document accordingly. ii. NONE This has already been addressed in the document accordingly. iii. NONE This has already been addressed in the document accordingly.				
No. Recommended	 This has been addressed in the document accordingly. ii. NONE This has already been addressed in the document accordingly. iii. NONE This has already been addressed in the document accordingly. iv. NONE This has already been addressed in the document 				
No. Recommended	 This has been addressed in the document accordingly. ii. NONE This has already been addressed in the document accordingly. iii. NONE This has already been addressed in the document accordingly. iv. NONE This has already been addressed in the document accordingly. 				

	This has already been addressed in the document accordingly and will not be moved into the section 1.1 in the document.
d.	ACCEPT
	This has been updated in the document accordingly.
e.	ACCEPT
	This has been mentioned in the document where appropriate.
f.	NONE
	It is best practice to have only a limited capability for editing programs and all amendments will need to be revalidated and so this will remain in the document as it is useful for the users to know.
g.	ACCEPT
	This has been updated in the document accordingly.
h.	ACCEPT
	This has been updated in the document accordingly.
i.	i. ACCEPT
	This has been updated in the document accordingly.
	ii. NONE
	The information will remain as it is in the document.
j.	ACCEPT This has been updated in the document.
k.	NONE
	This has already been removed from the document.

Comment nur	nber 6			
Date received	24/08/2017	Professional body	Society for Applied Microbiology	
Section	a. Introdu	a. Introduction		
	b. 1.1 Org	b. 1.1 Organisation of work		
	c. 2 Spec	c. 2 Specimen Processing		
	d. 5 Othe	d. 5 Other considerations to avoid contamination		
	e. 6 Qual	ality Assurance		
Comment	i			
a.				
-	• • •	uph 2] In addition to clones DNA and virus cell cultures, microbes e environment are a significant source of contamination.		

	ii.	[Paragraph 4] Another significant, or possibly greater, risk is through cross- contamination of different reactions prepared at the same time. Also, the contamination of master stocks (eg oligonucleotide stocks) by DNA templates is a major threat to be considered.
b.		
	i.	[Paragraph 2] The setup of a formal induction process should be a must, rather than a recommendation, to ensure all workers have a standardised introduction to a particular laboratory, regardless of prior experience.
	ii.	[Paragraph 3] We would recommend that batch numbers be recorded in a centralised manner for the laboratory, to improve traceability.
	iii.	[Paragraph 4] Some key examples of when gloves ought to be changed would be beneficial, for example in between the processing of individual patient samples.
	iv.	[Paragraph 5] Many laboratories are not able to arrange PCR work across separate rooms and therefore must rely on segregation of areas to separate pre- and post-PCR work. This means that changing laboratory coats between work areas is therefore not practical.
	v.	[Paragraph 7] Should also be clear here that when a fresh reagent arrives, it should be aliquoted into the amounts required for single use straight away. This is also relevant to section 2.3.
	vi.	[Paragraph 8] We would recommend that benches be wiped with disinfection solution before and after each procedure as necessary.
C.		
	i.	2.2 The unidirectional workflow: [Paragraph 3] It is also worth considering how air pressure and flow in these rooms/areas can be adjusted to minimise contamination risk. For example, the amplification room should be under negative air pressure to prevent PCR products potentially escaping and contaminating other rooms/areas of the laboratory. 2.3
	ii.	Reagent preparation clean room: [Paragraph 2] It may be necessary to clean the workspace when changing between different primers and other reagents.
	iii.	2.4. The nucleic acid extraction room [Paragraph 2] 'cDNA' should not be used here, as it may be misinterpreted as complementary DNA, which is not referred to in this section.
d.		
	i.	[Paragraph 1] For Real Time PCR, these glycosylases can show both advantages and limitations which ought to be recognised. Total elimination of contaminants is not always accomplished using this technique, particularly where PCR product length is short. Also inclusion of UDGs may reduce amplification efficiency and thereby delay or prevent detection, when only one or a few target molecules are present. Heat-labile forms of the enzyme are available to minimise residual UDG activity after PCR.
	ii.	[Paragraph 2] For consistency, the concentration of sodium hypochlorite should be stated.
	iii.	[Paragraph 3] It may be worth explicitly highlighting the risk of pipette contamination through not using aerosol-barrier tips. In addition, although

perhaps obvious, it could be specified that DNAse- and RNAse-free pipette tips, which have been gamma irradiated, should be utilised.

e. [Paragraph 1] We would advise that 'deep clean' decontamination procedures be put in place and employed in the event that a major lab disruption takes place (for instance during emergency evacuations or the entrance of building contractors into the PCR suite).

Into the PCR suite).		
Evidence		
Not completed.		
Financial barriers		
Not completed.		
Health benefits		
Not completed.		
Recommended	a. i. ACCEPT	
action	This has been updated in the document.	
	ii. ACCEPT	
	This has been updated in the document.	
	b. i. ACCEPT	
	This has been updated in the document.	
	ii. ACCEPT	
	This has been updated in the document.	
	iii. ACCEPT	
	This has been updated in the document.	
	iv. NONE	
	This will remain as described in the document. It may be necessary to review as part of a local risk assessment for the process.	
	V. ACCEPT	
	This has been updated in the document.	
	vi. ACCEPT	
	This has been updated in the document.	
	c. i. NONE	
	This is outside the remit of the UK SMIs.	
	іі. АССЕРТ	
	This has been updated in the document.	
	iii. ACCEPT	
	This has been removed from the document.	
	d. i. ACCEPT	

	This has been updated in the document accordingly.
	ii. ACCEPT
	This will be updated in the document.
	iii. ACCEPT
	This has been updated in the document accordingly.
e.	ACCEPT
	This has been updated in the document accordingly.

Comment number	7		
Date received	25/08/2017	Professional body	Royal Cornwall Hospitals Trust
Section	a. 5		
	b. 6		
Comment			
a. Section 5, 2nd bullet point - it would be useful to know the concentration of sodium hypochlorite required. Also is it better to state HCI as hydrochloric acid?			

b. Section 6, 3rd paragraph - 'characterized' (English or American spelling?)There is no mention of commercial reagents such as 'DNA away' for decontamination.

Evidence

Not completed.

Financial barriers

No.

Health benefits		
No.		
Recommended	a. ACCEPT	
action	This will be updated in the document.	
	b. NONE	
	This has already been updated in the document.	

Second consultation: 05/09/2017 - 19/09/2017

Version of document consulted on: Q 4dw+

Proposal for changes

Comment number 1

Date received	05/09/2017	Lab name	Sheffield Teaching Hospital NHS Trust		
Section	Multiple section	ons			
Comment					
Not all molecular assays are PCR. Either specify the UK SMI as PCR assays good laboratory practice or stop using PCR and replace with molecular amplification assays eg SDA, TMA etc.					
Evidence					
Not completed.					
Financial barriers					
Not completed.					
Health benefits					
Not completed.					
Recommended	NONE				
action			s it is and other molecular uded in the introduction.		

Comment number	2			
Date received	05/09/2017	Lab name	PHE NIS Porton	
Section	Several			
Comment				
Minor typo and three su	iggested additio	ns - see tracked-changed vers	sion upload.	
Evidence				
We have used previous experience of assay development and research to provide these comments which I hope are helpful.				
Financial barriers				
No.				
Health benefits				
No.				
Recommended	NONE			
action		were uploaded or received fro attempts requesting their comr		

Comment number	3				
Date received	06/09/2017	Professional body	Society for Applied Microbiology		
Section	General				
Comment					
We have an updated le	We have an updated logo, which is attached.				
Evidence					
Not completed.					
Financial barriers					
Not completed.					
Health benefits					
Not completed.					
Recommended	ACCEPT				
action	This will be amended accordingly in all the UK SMI templates.				

Comment number	4		
Date received	13/09/2017	Lab name	PHE
Section	General		

Comment

- a. General comment on the style and readability of the document: This UK SMI contains a great deal of valuable advice and information. It would benefit however, from being edited into a single style. For example, in paragraphs 3 & 4 of section 1.1 there is a mixture of the active and passive; please select one or the other. Similarly, must and should are used almost interchangeably but in para 3 of section 2.1 a couple of instances of 'shall' creep in. Please ensure the appropriate meaning has been used.
- b. Terminology: In section 2.2, reference is made to 'PCR workstation laminar flow cabinet'. Please note that such cabinets may not all use laminar flow; some use HEPA-filtered air (non-laminar), while others use still air.
- c. In section 6 where UK NEQAS is referred to, the word 'assurance' should be 'assessment' (according to the UK NEQAS web site).
- d. Typos etc.: Section 1.1, para 3. A 'policy' should be followed or observed; a 'procedure' or 'code of practice' may be practised.
- e. Section 2.1. Suggest 'However, this should not be into clean areas.'
- f. Section 5, para 7. This is a statement only and needs to be re-phrased as a recommendation.

g.	Quality management points:- Section 4, para 5. Where it is suggested that
	positive control material can be contrived specimens, the issue of commutability
	of the material could be included.

- h. Section 5, para 8. In this reminder about the importance of document control two key points have been included (up to date and current) but the need for management approval/authorisation has been omitted.
- i. Section 6, para 3. Assay validation is mentioned but equipment validation is not. That point could be inserted in section 5 para 5, to strengthen the importance of asset management.
- j. Reagent management. In light of an imminent PHE 'lessons learnt' report on the quality of reagents in molecular assays, there is an opportunity to include some additional best practice points about selection of suppliers, more detail on adequacy of acceptance testing procedures, and separately emphasis on risk management principles (in relation to ISO 15189) and also, perhaps, the value of end-to-end bar coding of samples and electronic reporting of results.

Evidence			
Not completed.			
Financial barriers			
No.			
Health benefits			
No.			
Recommended	a. ACCEPT		
action	This has been updated in the document accordingly. b. ACCEPT		
	This has been updated in the document accordingly. c. ACCEPT		
	This has been updated in the document accordingly.		
	d. NONE		
	This sentence will remain in the document as good housekeeping should be observed at all times in the laboratories.		
	e. ACCEPT		
	This has been updated in the document accordingly.		
	f. ACCEPT		
	This has already been updated in the document accordingly.		
	g. ACCEPT		
	This has been updated in the document accordingly.		
	h. ACCEPT		
	This has been updated in the document accordingly.		

i.	NONE
	Equipment validation is already in the UK SMI Q1 – "Evaluations, validations and verifications of diagnostic tests" document which is linked in the paragraph of Section 6. This has also been briefly mentioned in paragraph 5 of Section 5 that equipment used should be calibrated periodically.
j.	NONE
	This is outside the remit of this UK SMI document.

	_					
Comment number	5					
Date received	13/09/2017	Lab name	North West London Pathology			
Section	Title					
Comment						
	etitling the SMI	s no mention of other molecula - good practice when performi				
Evidence						
Scope of document						
This UK SMI describes key elements of how to organise facilities for molecular amplification assays. Introduction The ability of PCR to produce large numbers of copies of a target sequence from minute quantities - sometimes single copies - of DNA has provided the exquisite sensitivity that makes PCR a powerful diagnostic tool. However, this ability also necessitates that extreme care needs to be taken to avoid the generation of false positive results.						
Financial barriers						
None.						
Health benefits						
No.						
Recommended	NONE					
action	There is mention of Next Generation Sequencing technology.					

Comment number	6		
Date received	15/09/2017	Lab name	Royal Cornwall Hospital
Section	Several		
Comment			

We discussed this UK SMI with staff who do and don't work in Molecular.

- a. The novices asked if more information could be provided in the Introduction (e.g. the word PCR what does it stand for; what is the difference between molecular and PCR testing?)
- b. Section 5, pg14 perhaps more explanation on good pipette technique (ie smooth, downward pressure to reduce contamination/aerosols).
- c. Also what is 'regular' environmental swabbing how often?
- d. No mention of having separate fridges in each room. If there reagents are kept in a fridge in the pre-PCR room and they are taken into another room, they cannot be returned.

Evidence

Not completed.

Financial barriers

No.

Health benefits

No.

110.	
Recommended action	a. ACCEPT
	The acronym PCR has been written in full.
	b. NONE
	This is not within the remit of this UK SMI document.
	c. NONE
	The frequency of how environmental swabbing of areas where high throughput PCRs are performed is down to local policies. However, it is advisable to do so to avoid contamination.
	d. NONE
	This is not within the remit of this UK SMI document.

Respondents indicating they were happy with the contents of the document

Overall number of comments: 5			
Date received	16/08/2017	Lab name	Virology, The James Cook University Hospital, Middlesbrough
Health benefits	·		

Ű,	h themselves and	have health benefits for staff d their colleagues from unner	0
Date received	18/08/2017	Lab name	Senior clinical scientist
Health benefits			
No.			
Date received	21/08/2017	Lab name	Animal and Plant Health Agency
Health benefits			
No, as staff who work w	ith zoonotic disea	ases will be on occupational h	nealth schemes.
Date received	08/09/2017	Lab name	Antrim Area Hospital Microbiology Laboratory
Health benefits			
No.			
Date received	12/09/2017	Lab name	Sheffield Teaching Hospitals Microbiology
Health benefits	1		
No.			